

Test set "Measuring-Testing-Checking" for DIN EN ISO 9934

with radiation meter KD 420 and field strength meter KD 400

In carrying case, made from plastic, with following contains:

- Field strength meter KD 400 with calibration-standard
- Digital UV radiation and daylight measuring device KD 420 with measuring head UV/VIS DIN EN ISO 9934-2
- Test piece in accord Prof. Berthold
- Reference Test Piece Type 1 accord. to DIN EN ISO 9934-2
- ASTM centrifuge tube with stand
- Magnetoscope

Order Number: 1400143



Field strength meter „KD 400“ Including: tangential field probe, carrying case. The technical device measures D.C. and A.C. magnetic fields, values are shown on LCD display (3 ½ digits) with an accuracy +/- 2 %. Switchable between A/cm, Milli-Tesla, Gauß, Oerstedt
Measuring range: 0 – 400 A/cm

Order number: 1400130, 1400129



Radiation meter KD 420 for measuring daylight (white light component) and for measuring the UV/A/B/C

Order Number: 1400200, 1400201



Magnetoscope
The technical device measures residual magnetism.

Measuring range for residual magnetism:
0 – (+)20 Gauß
0 – (-)20 Gauß

Order Number: 1400104




ASTM centrifuge with stand
For control concentration of test suspension (especially new preparation) .Sedimentation glass with fine scaling.

Order Number: 1400103



Test Piece according Prof. Berthold
The technical device defines direction of magnetic field and shows incorporated artificial errors.

Order Number: 1400101



Reference Test Piece 1
With photo for comparison and manufacturer certificate to control sensitivity of test fluid.

According DIN EN ISO 9934-2

Order Number: 1400102